

Certificate of Analysis

Product: Water
Item Number: WAD
Grade: Lab
Lot Number: 8282023
Manufacture Date: 08/21/2023
Expiration Date: 08/20/2025
Country of Origin: United States

| Tested Property | Specification | Analysis |
|--|-----------------|----------|
| Chloride(Cl) | max 0.4 ppm | Conforms |
| Conductivity at 25°C | max 0.056 µS/cm | Conforms |
| Heavy Metals (as Pb) | max 0.01 ppm | Conforms |
| LC Suitability (Absorbance) | To Pass Test | Conforms |
| LC Suitability (Gradient Elution Test) | To Pass Test | Conforms |
| Nitrate (NO ₃) | max 0.4 ppm | Conforms |
| Organic Carbon (TOC) | max 50 ppb | Conforms |
| Oxidizable Substances (Permanganate Retention) | To Pass Test | Conforms |
| Phosphate (PO ₄) | max 1.0 ppm | Conforms |
| Silicate (as SiO ₂) | max 0.01 ppm | Conforms |
| Sulfate (SO ₄) | max 1.0 ppm | Conforms |

Water, Reagent, Suitable for Ultraviolet Spectrophotometry and Liquid Chromatography

Water

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley