

Certificate of Analysis

Product: Sodium Thiosulfate Pentahydrate

Item Number: C7810

Grade: ACS

Lot Number: 240206-A

Manufacture Date: 05/02/2023

Expiration Date: 05/02/2028

Country of Origin: China

Tested Property	Specification	Analysis
Assay	99.5 - 101.0%	99.9%
pH of a 5% solution @ 25°C	6.0-8.4	Passes
Insoluble Matter	≤0.005%	<0.005%
Nitrogen Compounds	≤0.002%	<0.002%
Sulfate & Sulfite	<0.1%	0.09%
Sulfide	To Pass Test	Passes

Sodium Thiosulfate Pentahydrate

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com.

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley