

Certificate of Analysis

Product: Water Purified
Item Number: WAU
Grade: USP
Lot Number: 5000795/1.1
Manufacture Date: 08/21/2024
Expiration Date: 08/20/2025
Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Clear, colorless liquid	Passes
Water Conductivity	$\leq 1.3 \mu\text{S/cm}$ at 25°C	0.18
Microbial Limits	≤ 100 CFU/mL	<1
pH	5.0 - 7.0	6.3
Substances reducing Permanganate	The red color doesn't disappear	Pass
Total Organic Carbon	<0.50 mg C/L	0.15

[Water Purified](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.