



Certificate of Analysis

Product: Hydrogen Peroxide 30% Solution

Item Number: HPA30

Grade: ACS/Reagent

Lot Number: 220504

Manufacture Date: 05/04/2022

Expiration Date: 05/03/2024

Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Colorless and free from suspended matter or sediment	Passes
Assay, %	29 - 32	31.3
Color (APHA)	Max. 10	5
Titratable Acid, meq/g	Max. 0.0006	<0.0006
Chloride, ppm	Max. 2	<1
Nitrate, ppm	Max. 2	<1
Phosphate, ppm	Max. 2	<1
Sulfate, ppm	Max. 5	<0.5
Ammonium, ppm	Max. 5	<1
Heavy Metals as Pb, ppm	Max. 1	<0.1
Iron, ppm	Max. 0.5	<0.1
Sodium Stannate, ppb	Max. 500	<300

This product is slightly stabilized with Sodium Stannate below 500 ppb for safety purposes.

[Hydrogen Peroxide 30% Solution](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for

Phone: 512-668-9918, Fax: 512-886-4008,

E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley



LAB ALLEY
ESSENTIAL CHEMICALS



particular purposes are made without warranty, expressed or implied.
This document was electronically issued and is therefore valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.
© copyright: 2023 Lab Alley