

Certificate of Analysis

Product: Water
Item Number: WAHPLC
Grade: HPLC
Lot Number: 230220-B048867
Manufacture Date: 02/17/2023
Expiration Date: 02/16/2026
Country of Origin: United States

Tested Property	Specification	Analysis
Optical Absorbance @ 190 nm	Max 0.01	0.00
Optical Absorbance @ 200 nm	Max 0.01	0.00
Optical Absorbance @ 210 nm	Max 0.01	0.00
Optical Absorbance @ 250 nm	Max 0.01	0.00
Specific Conductance at 25°C	Max 1.5 µS/cm	0.36
Fluorescence emis. for max impurities @ 450nm	Max 0.1 ppb	<0.1
LC Gradient Elution Suitability at 254 nm	To pass test	Passes
Substances reducing Permanganate	To pass test	Passes
Residue after evaporation	Max 1 ppm	0

Water

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.