

Certificate of Analysis

Product: Ethanol 200 Proof (100%) Completely Denatured with Heptane

Item Number: EAC200

Grade: Reagent

Lot Number: 5000206/1.1

Manufacture Date: 07/17/2023

Expiration Date: 07/16/2025

Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Clear thin liquid	Pass
Color (APHA)	Max 15 Pt-Co	<15
Odor	Characteristic	Pass
Suspended matter	Substantially free from particulate	Pass
Specific gravity @ 60°F	0.785 - 0.793	0.789
Water Content, % (ppm)	<0.2 (2000)	0.05 (460)
Identity by GC	Compared to standard	Pass
Ethanol, Wt %	Typical Composition	95.8%
n-Heptane, Wt%	Typical Composition	4.2%

[Ethanol 200 Proof \(100%\) Completely Denatured with Heptane](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com.

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley