

Certificate of Analysis

Product: Acetonitrile
Item Number: CANLCMS
Grade: LCMS
Lot Number: 5000280/1.1
Manufacture Date: 11/04/2022
Expiration Date: 11/03/2025
Country of Origin: United States

Tested Property	Specification	Analysis
Assay (by GC-FID)	NLT 99.9%	99.99
Color Pt-Co	Max 5	5
Fluorescence at emission for Impurities	Max.1 ppb @ 254nm	LT 1 ppb
GC-ECD Responsive residue	Max. 10 ppb as Heptachlorepoxyde	LT 1 ppb
GC-FID Responsive residue (single peak)	Max. 5 ppb as 2-Octanol	LT 1 ppb
LC gradient elution suitability at 210 nm	Max. 5 mAU	1.0
LC gradient elution suitability at 254 nm	Max. 5 mAU	0.5
Potassium (K)	Max. 0.1 ppm	0.001
Silver (Ag)	Max. 0.1 ppm	LT 0.001 ppm
Aluminum (Al)	Max. 0.5 ppm	LT 0.001 ppm
Barium (Ba)	Max. 0.1 ppm	LT 0.001 ppm
Calcium (Ca)	Max. 0.1 ppm	0.001 ppm
Cadmium (Cd)	Max. 0.5 ppm	LT 0.001 ppm
Cobalt (Co)	Max. 0.02 ppm	LT 0.001 ppm
Chromium (Cr)	Max. 0.02 ppm	LT 0.001 ppm
Copper (Cu)	Max. 0.01 ppm	LT 0.001 ppm
Iron (Fe)	Max. 0.02 ppm	0.001
Magnesium (Mg)	Max. 0.01 ppm	LT 0.001 ppm
Manganese (Mn)	Max. 0.02 ppm	LT 0.001 ppm
Sodium (Na)	Max. 0.1 ppm	0.001 ppm

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com.

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley

Nickel (Ni)	Max. 0.5 ppm	LT 0.001 ppm
Lead (Pb)	Max. 0.1 ppm	LT 0.001 ppm
Tin (Sn)	Max. 0.1 ppm	LT 0.001 ppm
Zinc (Zn)	Max. 0.1 ppm	LT 0.001 ppm
LC-MS Negative Mode (as Nitrophenol)	Max. 50 ppb	LT 50 ppb
LC-MS Positive Mode (as Reserpine)	Max. 50 ppb	LT 50 ppb
UV Absorbance @190nm	Max. 1.00 AU	0.45
UV Absorbance @195nm	Max. 0.14 AU	0.05
UV Absorbance @200nm	Max. 0.05 AU	0.01
UV Absorbance @210nm	Max. 0.04 AU	0.01
UV Absorbance @220nm	Max. 0.02 AU	0.01
UV Absorbance @230nm	Max. 0.01AU	0.003
UV Absorbance @254nm	Max. 0.005 AU	0.001
UV Absorbance @400nm	Max. 0.005 AU	0.001

Acetonitrile

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.