

Certificate of Analysis

Product: Dichloromethane, 99%
Item Number: DCMHPLC
Grade: HPLC
Lot Number: 220831
Manufacture Date: 08/31/2022
Expiration Date: 08/30/2026
Country of Origin: China

Tested Property	Specification	Analysis
Assay	Min 99.8 wt%	99.99
Residue on evaporation	Max 0.0005 wt%	0.0002
Water by KF	Max 0.02 wt%	0.0036
Titration Acid	Max 0.0003 mmol/g	0.00002
Density (g/ml) @ 20°C	1.320 - 1.330	1.326
UV Absorbance (1 cm cell vs water)		
@ 240 nm	Max 0.4	0.105
@ 250 nm	Max 0.1	0.014
@ 260 nm	Max 0.05	0.006

[Dichloromethane, 99%](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This document was electronically issued and is therefore valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley