



Certificate of Analysis

Product: Xylenes Histological Lab Grade
Item Number: XYL
Grade: Lab
Lot Number: 0084L01044
Manufacture Date: 11/22/2021
Expiration Date: 11/21/2024
Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Clear colorless liquid	Pass
Assay (Xylene Isomers+Ethylene Benzene)	>98.50%	99.69%
Meta-Xylene	As reported	47.47%
Para-Xylene	As reported	21.44%
Ortho-Xylene	As reported	13.86%
Ethyl Benzene	<25.0%	16.91%
Color, Pt-Co	Max 20	5
Toluene	Max 0.5%	0.0283%
Sulfur	Max 1.0 mg/kg	<1.0
Water	Max 0.05%	0.01%
Distillation Range IBP	Min 137 °F	138 °F
Distillation range DP	Max 143 °F	139.8 °F
Acidity	As reported	No free acid

[Xylenes Histological Lab Grade](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley



LAB ALLEY
ESSENTIAL CHEMICALS



This document was electronically issued and is therefore valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: Customerservice@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2023 Lab Alley